

PC-Display Version History

- 1.53__
Added shading under CO2 and respiration waveforms
Changed default alarm limits
- 1.52__
Expanded Breed to full number of characters for Avimark interface
- 1.51__
Complete re-write of Nag screen and file recording code.
- 1.50__
Avimark interface for non Banfield users
- 1.49__
Fixed file naming bugs when patient ID and last name are changed
- 1.48__
Fixed Banfield Nag screen recursion bug
- 1.47__
Fixed New patient via disconnect bugs (Banfield)
Fixed DEMO recording bug
- 1.46__
Cleaned up file save bug (Banfield)
Reduced Nag screen trip criteria to Patient Name and Last Name
Reduced width of Vet information fields
Added message box to service terminal flash update routine.
- 1.45__
Fixed some more nag screen bug in Banfield version
Fixed alarm bug introduced by nag screen
Added Dental Scalar rejection mode.
- 1.44__
Fixed more nag screen bugs in Banfield version
Fixed Apnea alarm bug
Changed connection thread to require patient information prior to attempting a connection
Changed default CO2 inspired low alarm to be disabled
- 1.42__
Added Monitoring Guidelines to Help menu
Added BP Error Messages
- 1.41__
Added Bluetooth compatibility with Windows XP SP3
Added Monitoring Guidelines to help screen
Updated manual.
Fully implemented Banfield nag screen.
- 1.40__
Fixed Banfield nag screen bug caused when closing out application while recording with no Patient data entry

- 1.39__
 - Darkened CO2 trace on printer output
 - Added remote menu for cuff size selection when BP measurement is started
 - Updated Manual
 - Fixed configuration bug with VetGard I
 - Added conformation message box after zeroing command is sent to gas probe
 - Added nag screen (Banfield only) to ask for Patient data entry at exit

- 1.38__
 - Changed border color on main display from white to gray
 - Updated Remote Switcher for non XP/SP2 users
 - Changed CO2 print waveform width and color

- 1.37__
 - Added ability to save CO2 waveform to .pve file
 - Added CO2 save configuration switch
 - Added Demo watermark for display and printout from PC-Display
 - Changed patient data entry dialog to birth date or age
 - Added a field to patient data entry for first/last name
 - Added hospital info setup in configuration for Banfield
 - Changed filenames to include patient_clientlastname_hospitalnum_xxxx.pve
 - Fixed tab selection on patient data entry
 - Increased status display size of the date, time and continuous ECG counter
 - Fixed reconnect bug
 - Fixed Banfield N2O bug
 - Fixed Banfield CO2 tab bug

- 1.33__
 - Added Import and Export for settings
 - Added pull down menus and customization for patient data entry
 - Changed filenames to include Patient_Owner_Clinic_xxxx.pve
 - Added EULA to install

- 1.32__
 - Added breakout and clarification of CO2 probe error messages.
 - Decreased PC-Vet com message processing interval
 - Fixed PC-Vet leads off bug introduced in v1.31

- 1.31__
 - Fixed single instance capability in Edit Default Config.
 - CO2 reorder message breakout to CO2_Reorder.txt input file
 - Fixed PC-Vet manufacturing terminal bug
 - Fixed mouse over icon ID bug
 - Added Support_URL.txt customization.

Banfield Build changes

 - Added separate Banfield Build
 - Added Banfield Configuration Lockouts for Defaults and Vet Info
 - Eliminated splash screen for Banfield build
 - Vmed, VetGard+ names swapped for Banfield, PetProtect

- 1.30__
 - Added Manufacturing over ride for password to get into service terminal

- 1.29__
 - Added Watermark for Demo report printouts SYS-3008
 - Fixed intermittent exception bug when printing

- Added Banfield "AutoSave" feature SYS-3003
- Layed in groundwork for "CO2 save" option.
- Added Banfield "Single Instance" feature SYS-3011
- Disabled Keyhits from remote from working while in Service Terminal for mfg test.
- Updated online manual.

1.28

- Resolved reported issue with random numbers in episode vitals
- Fixed missing ECG beeps when PC-Display is minimized
- Resolved alarm issues

1.26

- Added in Phase-In suggestions from host validation
- Changed to MSI installer

1.25

- Reduced CPU load significantly

1.23

- Fixed file corruption bug on recorded ECG .pve files

1.22

- Added ability for non-MS bluetooth users to connect to multiple remotes without changing congig settings

1.21

- Fixed bug that caused timeouts with remotes so that users would frequently lose connection. This also fixed many problems with blood pressure not working.
- Fixed bug where the respiration rate from the CO2 probe would not go below 10 breaths per minute.
- Went to a simpler version numbering scheme.

.
. .
.

1.0.14

- Added the inclusion of gas and blood pressure vitals information to the patient episodes recorded while connected to a VetGard+ remote. Consequently, the printed events now take up 4 lines on a page just so they will all fit.

.
.

1.0.6

- Official release.